

Page 1 of 2

RWP/Rev. . . .
TF-100/Rev:012

Description of Work/Comments:

241 -AY Annulus Emf calibration and functional tests. Specifically AY-102-152. Dose Rates were taken with the displacer at the sight glass and again at the pipe below the first flange. Area in AY Farm for this evolution is posted RA/CA. Rad instruments were released to Radiological Protection and Instrument Tech Instruments were released to the Instrument Technicians.

100

BD-6003-343 (REV 4)

TANK OPERATIONS CONTRACTOR RADIOLOGICAL SURVEY REPORT

RSR No. WTO-022526

Page 2 of 2

Map/Sketch

N/A

Legend:

☒ Direct Measurement☒ Air Sample☒ Smear☒ LAW☒ NeutronField Dose Rate
* Contact Dose RateNote: Dose Rates in mrem/hr
unless otherwise noted.

----- (designation inside) -----

Instruments

Instrument/Probe Model	RO-20	GM Model g-m Probe	PAM Probe	N/A	N/A	N/A	N/A	N/A	N/A
Serial No.	ICEB4-0071	CMEB3-0160	ACBC1-0032	N/A	N/A	N/A	N/A	N/A	N/A
Efficiency (used)	N/A	DTHNC-0931	DTHN3-0026	N/A	N/A	N/A	N/A	N/A	N/A
Air Sample Log Number		Smear Log Number							
1	N/A	<input type="checkbox"/> Breathing Zone <input type="checkbox"/> General Area	N/A						
2	N/A	<input type="checkbox"/> Breathing Zone <input type="checkbox"/> General Area	N/A						
HPT: Date: 10-12-2011				Payroll No: 8D399		Initial Review Date & Initials: 10/13/11			
Name (Print): J. Schaeffer				Name (Print):		Final Review Date & Payroll No.: 10/19/11			
Signature: <i>J. Schaeffer</i>				Signature: <i>Greg Levy</i>		20537			